List of Tables

2.1	Sensor types supported by the different variant smartphones \ldots .	37
2.2	Detailed specifications of the Sony Xperia E3 smartphone	39
3.1	Accuracy and precision data for concentrations of standard fluoride	
	$solutions(n = 5) \dots $	70
3.2	Comparison of the smartphone based photometric platform with	
	the existing portable spectrophotometers	74
3.3	Accuracy and precision data for concentrations of standard solu-	
	tions of iron(II) and phosphate (n=5)	89
5.1	Some important technical specification of the POF \ldots	123
6.1	Detailed specifications of the iPhone 4 smartphone, apple Inc. $\ . \ .$	143
6.2	Detailed specifications of the Samsung Galaxy J7 prime smartphone	150

xxii

List of Figures

- 1.1 Estimated compound annual growth rate in mobile-broadband subscription globally. (Source: ICT fact and figure 2017 by ITU). . . . 4
- 1.2 Smartphone based colorimetric detection methods; (a) working procedure of the feature phone camera based colorimetric sensor, reproduced from [23] with permission from American Chemical Society, (b) first demonstration of smartphone based standalone colorimetric platform for detection of cancer biomarker in urine, reproduced from [24] with permission from Royal Society of Chemistry, (c) demonstration of smartphone based colorimetric platform with 3D printed attachment, reproduced from [26] with permission from Royal Society of Chemistry, and (d) reference based colorimetric platform, reproduced from [28] with permission from Elsevier. . . .
- 1.3 The development of smartphone based spectroscopic methods; (a) working procedure of an absorption based smartphone spectrometer, reproduced from [36] with permission from Royal Society of Chemistry, (b) demonstration of smartphone based fluorescent spectrometer, reproduced from [37] with permission from American Chemical Society and (c) demonstration of smartphone based reflectance spectroscopy, reproduced from [38] with permission from Royal Society of Chemistry.
- 1.4 The development of smartphone based microscopic imaging platforms; (a) demonstration of mobile microscopy platform by associating traditional optical component, reproduced from [46] with permission from Public Library of Science, (b) cellphone based lensfree microscopy platform, reproduced from [47] with permission from Royal Society of Chemistry, and (c) demonstration of smartphone based fluorescent microscopy, reproduced from [56] with permission from American Chemical Society.

7

11

9

- 1.5 The development of smartphone ALS based sensing platforms; (a) working of a smartphone ALS based biosensor where ambient light is used as a light source [62] with permission from Elsevier, (b) demonstration of smartphone ALS based plasmonic nanosensor read-out device, reproduced from [63] with permission from Royal Society of Chemistry and (c) demonstration of smartphone ALS as a platform for educational experiments, reproduced from [65] with permission from Institute of Physics.
- 1.6 The development of smartphone NFC based sensing platforms; (a) working procedure of the NFC based wireless gas sensing scheme, reproduced from [69] with permission from National Academy of Sciences (United States) and (b) demonstration of NFC based smart bandage for wound monitoring, reproduced from [71] with permission from Elsevier.
- 1.8 The development of smartphone audio jack based sensing platforms; (a) Demonstration of embedded audio jack of the smartphone as a potentiostatic control for ECL based applications, reproduced from [77] with permission from Elsevier, (b) audio jack based universal mobile electrochemical detector for POC applications in resource poor settings, reproduced from [78] with permission from American Association for the Advancement of Science, and (c) audio jack based potentiostat for cyclic voltametric applications in a smartphone platform, reproduced from [80] with permission from Else-

. 17

13

vier.

.

2.1	Embedded ALS in a Sony Xperia E3 smartphone: (a) position of	
	the ALS, (b) photo image of integrated ambient light and proximity	
	sensor chip, and (c) response characteristic of the photodiode chan-	
	nels in Avago APDS-9930 sensor chip as adopted from the datasheet	
	[7]	38
2.2	Functional block diagram of Avago APDS-9930 sensor chip as adopted	
	from the datasheet [7]. \ldots \ldots \ldots \ldots \ldots \ldots	39
2.3	(a) Optical set-up for evaluation of ALS performance with that of	
	a laboratory grade photo-detector and (b) Response characteristics	
	for both the detectors	41
2.4	Optics design of a smartphone camera.	42
2.5	Schematic of Bayer pattern generation and digital colour image	
	detection by a smartphone CMOS sensor.	43
2.6	Emission spectra of the LED flash embedded in the Sony Xperia	
	E3 smartphone recorded by optical sprectrometer.	44
2.7	Circuit diagram for powering an external LED by the smartphone	
	battery.	44
2.8	Functional block diagram of Avago APDS-9930 sensor chip as adopted	
	from the datasheet [7]. \ldots \ldots \ldots \ldots \ldots \ldots	45
2.9	MIT App inventor Designer webpage.	46
2.10	Components in Designer webpage for designing user interface of the	
	application.	47
2.11	MIT App inventor Block editor webpage.	49
2.12	Examples of MIT App inventor block codes	50
2.13	Block code used to determine light intensity and a screenshot of the	
	developed application.	50
3.1	Photo image of the prepared samples that contain different concen-	
	tration of fluoride level. \ldots \ldots \ldots \ldots \ldots \ldots \ldots \ldots	58
3.2	Absorption spectrum of zirconium $-{\rm SPADNS}$ dye mixed with water	
	sample containing fluoride ion at different concentrations. $\ . \ . \ .$	58
3.3	schematic view of the optical configuration for wavelength selection	
	from the LED flash of the smartphone.	59

schematic view of the compact optical attachment used for light	
coupling and sample handling.	60
(a) Schematic of the smartphone based photometric platform and	
(b) photograph of the designed platform	61
Process flow of the custom designed smartphone application. $\ . \ .$	63
Screenshot of the block codes developed with the MIT app inventor	
interface for implementing regression analysis	65
Response characteristics of the designed smartphone based photo-	
metric platform. (a) response curve for the range 0 mg/L to 3 mg/L	
and (b) Calibration curve prepared for 0 mg/L to 2 mg/L	67
Response characteristic of the designed photometric platform for	
standard samples with a lower level of fluoride concentrations in	
the range 0 mg/L to 0.1 mg/L. \ldots	69
Fluoride concentration measurement for field collected water sam-	
ples. (a) Histogram representation of the fluoride concentration es-	
timation by a commercial spectrophotometer and by the designed	
sensor. (b) A screenshot image of one of the considered areas (Ham-	
ren, Karbianglong) shown in a Google maps application, Google	
Inc. (used with permission), to demonstrate the in-field data shar-	
ing ability of the designed sensor	71
Response characteristics of the designed photometric platform for	
two different smartphones.	73
Photo image of the prepared iron (II) standard solutions. \ldots .	77
Photo image of the prepared phosphate standard solutions. \ldots	79
(a) Absorption spectra of the reagent treated standard iron(II) and	
(b) phosphate solutions. \ldots \ldots \ldots \ldots \ldots \ldots \ldots	79
(a) Schematic of the smartphone based photometric platform and	
(b) photograph of the designed platform. \hdots	81
Workflow of the designed $\mathbf{PhotoSense}$ smartphone application	82
Screenshot of Block codes developed with the MIT app inventor	
interface for averaging intensity data for 20 seconds. \ldots	83
. Screenshot of Block codes developed with the MIT app inventor	
interface for evaluation of absorbance value	85
	 coupling and sample handling. (a) Schematic of the smartphone based photometric platform and (b) photograph of the designed platform. Process flow of the custom designed smartphone application. Screenshot of the block codes developed with the MIT app inventor interface for implementing regression analysis. Response characteristics of the designed smartphone based photometric platform. (a) response curve for the range 0 mg/L to 3 mg/L and (b) Calibration curve prepared for 0 mg/L to 2 mg/L. Response characteristic of the designed photometric platform for standard samples with a lower level of fluoride concentrations in the range 0 mg/L to 0.1 mg/L. Fluoride concentration measurement for field collected water samples. (a) Histogram representation of the fluoride concentration estimation by a commercial spectrophotometer and by the designed sensor. (b) A screenshot image of one of the considered areas (Hamren, Karbianglong) shown in a Google maps application, Google Inc. (used with permission), to demonstrate the in-field data sharing ability of the designed sensor. Photo image of the prepared iron(II) standard solutions. Photo image of the prepared phosphate standard solutions. (a) Absorption spectra of the reagent treated standard iron(II) and (b) phosphate solutions. (a) Schematic of the esigned platform. (b) photograph of the designed platform. (c) Screenshot of the casigned platform. (d) Screenshot of the esigned platform. (e) Screenshot of the esigned platform. (f) photograph of the designed platform. (h) photograph of the designed platform.

3.19	Characteristic plot of absorbance measured with the sensor for the
	concentration range 0 mg/L to 5 mg/L of the prepared standard
	solutions
3.20	Characteristic plot of absorbance measured with the sensor for the
	concentration range 0.01 mg/L to 0.10 mg/L of the prepared stan-
	dard solutions.
3.21	Comparison of the designed sensor with a laboratory grade spec-
	trophotometer in measuring (a) iron (II) and (b) phosphate in water. 90
4.1	schematic of a basic nephelometer
4.2	Photo image of the prepared formazin standard
4.3	A schematic diagram of the proposed setup with different compo-
	nents (b) photo image of the designed smartphone turbidimeter (c)
	inside view of the turbidimeter compartment
4.4	workflow of the designed smartphone based application 105
4.5	Response characteristics of the designed smartphone based neph-
	elometric platform in the turbidity ranges (a) 0 NTU to 1.0 NTU
	(b) 1.0 NTU to 10 NTU, and 40 NTU to 400 NTU
4.6	Response characteristics of the designed smartphone based neph-
	elometric platform in the turbidity ranges 0 NTU to 1.0 NTU in
	terms of intensity units
4.7	The uncertainty of measurement found for 10 sets of measurement
	of turbidity of standard formazin samples in the turbidity range 0
	NTU to 1.0 NTU
4.8	Histogram representation of comparison of turbidity measurement
	of the field collected and laboratory prepared clay-mixed water sam-
	ples by the standard turbidity meter and by developed smartphone
	based nephelometric platform
4.9	Response characteristics of the designed smartphone based neph-
	elometric platform in measuring different colored turbid water sam-
	ples
5.1	Propagation direction of the evanescent wave in a dielectric inter-
	face

5.2	U-shaped configuration for enhancing sensitivity of fiber optic evanes-
	cent wave sensor
5.3	(a) Schematic of smartphone based fiber optic evanescent wave sen-
	sor and (b) photo image of the designed sensor
5.4	Screenshot images of the designed 'SalinitySense' application. $\ . \ . \ . \ 126$
5.5	Response characteristic of the smartphone based fiber optic evanes-
	cent wave sensor in the salinity range 0 ppt to 100 ppt
5.6	Response characteristic of the smartphone based fiber optic evanes-
	cent wave sensor in the salinity range (a) 0 ppt to 1.0 ppt, (b) 1
	ppt to 10 ppt, and 10 ppt to 100 ppt
5.7	Response characteristics of the designed smartphone based fiber
	optic evanescent wave sensor in the salinity ranges 0 ppt to 1.0 ppt
	in terms of intensity units
5.8	Uncertainty in measurement for both the sensing module for the
	salinity range 0 ppt to 1.0 ppt
5.9	Stability measurement data the designed sensor for solution with
	salinity level of 35 ppt
5.10	Repeatability characteristics of the designed sensor
5.11	Histogram representation of comparison of salinity measurement of
	prepared water samples by the standard conductivity meter and by
	the designed smartphone salinity sensor
6.1	Schematic representation of the interference phenomenon which oc-
	curs as a result of back-reflection of the incident light signal from a
	glass plate
6.2	(a) Schematic of the designed smartphone based interferometric
	system and (b) photo image of the designed system
6.3	(a) Photo image of the captured interferograms and (b) variation
	of intensity of the central fringe with incident angle
6.4	(a) Cropped images of the central fringe of the interferograms and
	(b) Characteristic plot of normalized intensity versus the angular
	rotations of the glass slide

6.5	Schematic of the optimized experimental set-up for smartphone
	based interferometric system
6.6	Schematic of the optimized experimental set-up for smartphone
	based interferometric system
6.7	Different opto-mechanical components printed to develop the 3D
	printed interferometric system
6.8	(a) 3D design of the smartphone based interferometric system and
	(b) photo image of developed system
6.9	Detailed workflow of the image processing algorithms implemented
	in the FringeApp application
6.10	(a) Screenshot images of the first window of the designed FringeApp
	application showing the reference and the interferogram with arbi-
	trary angle and (b) screenshot image of the detection window devel-
	oped for automatic fringe processing and evaluation of the angular
	rotation
6.11	Photo images of the captured interferogram for different OPD in-
	tervals
6.12	(a) variation of fringe order and (b) variation of distance of the
	fringe center line from the boundary of the interferogram with the
	angular rotation of the glass slide
6.13	(a) Cropped photo images of the central fringes when the glass
	slide is rotated with a step of 0.02° and (b) characteristic plot of
	normalized intensity versus the angular rotations of the glass slide
	in the first OPD interval
6.14	The characteristic plots of intensity versus angular rotation of the
	glass slide for different OPD intervals
6.15	Linear fitting of the intensity versus angular rotation of the glass
	slide for different OPD intervals

List of abbreviations

POC	Point of Care
LDC	Least Developed Country
ITU	International Telecommunication Union
CMOS	Complementary Metal-Oxide-Semiconductor
CCD	Charge-Coupled Device
CPU	Central Processing Unit
GPU	Graphical Processing Unit
GUI	Graphical User Interface
PC	Personal Computer
LED	Light Emitting Diode
ALS	Ambient Light Sensor
USB	Universal Serial Bus
NFC	Near Field Communication
IR	Infra Red
NIR	Near Infra Red
CFA	Color Filter Array
SPR	Surface Plasmon Resonance
LSPR	Localized Surface Plasmon Resonance
RFID	Radio-Frequency Identification
CARD	Chemically Actuated Resonant Device
SWCNT	Single-Walled Carbon Nanotubes
ECL	Electro-Chemiluminescence
ISO	International Organization for Standardization
OS	Operating Systems
ADC	Analog-to-Digital Converter

UV-VIS	Ultraviolet-Visible
SD	Standard Deviation
RSD	Residual Standard Deviation
ISE	Ion Selective Electrodes
NTU	Nephelometric Turbidity Units
USEPA	United States Environmental Protection Agency
IS	Indian Standards
JCGM	Joint Committee for Guides in Metrology
BP	Blood Pressure
EC	Electrical Conductivity
ATR	Attenuated Total Reflection
POF	Plastic Optical Fiber
DI	Deionized
OPD	Optical Path Difference
CAD	Computer-Aided Design
IDE	Integrated Development Environment

List of symbols

3G	Third Generation
4G	Fourth Generation
mg	Milligram
mL	Milliliter
mg	Milligram
nm	Nanometer
$\mu { m m}$	Micrometer
mA	Milliampere
R	Resistance, Ω
λ	Wavelength
A	Absorbance
ϵ_{λ}	Molar absorption coefficient
b	optical pathlength
C	Concentration
P_0	Incident radiant power
Р	Transmitted radiant power
T	Transmittance
S	Sensitivity
A.U.	Arbitrary Unit
R	Resolution
I_0	Incident light intensity
Ι	Transmitted light intensity
τ	Extinction coefficient
L	Scattering path length
N	Number of particles

a	Particle radius
π	Pi
Q	scattering coefficient
μ_p	Refractive index of the particle
μ_m	Refractive index of the medium
m	Refractive index ratio
α	Size parameter
k	Wave vector
θ	Scattering angle
R_{ALS}	Resolution of the ALS
σ_m	Standard deviation of the mean
θ_c	Critical angle
d_p	Penetration depth
γ	Evanescent absorption co-efficient
n	Refractive index
t	Thickness
m	Fringe order
$\delta { m m}$	Change in fringe order
f	Focal length