

DECLARATION

This is to certify that the thesis entitled “*Development of composite conducting polymer with 2D layered and metallic nanosystems for biofunctionalization and electrochemical biosensing applications*” being submitted to the School of Sciences, Tezpur University, Tezpur, Assam in partial fulfilment for the award of the degree of Doctor of Philosophy in Physics is a record of bona fide research work accomplished by me under the supervision and guidance of Prof. Dambarudhar Mohanta.

All help received from various sources has been duly acknowledged. No part of this thesis has been submitted elsewhere for the award of any other degree.

Date: 31/07/2025

Place: Tezpur

Susmita Baruah

Susmita Baruah

Department of Physics

School of Sciences

Tezpur University

Tezpur-784028, Assam

Dedication

*I Dedicate this thesis to my
beloved parents*



Tezpur University
(A Central University by an act of Parliament)
Department of Physics
Tezpur – 784028, ASSAM, INDIA

Prof. Dambarudhar Mohanta
Professor
Department of Physics
Tezpur University, Tezpur

Ph. +91-8876507550 (M)
Email: best@tezu.ernet.in

CERTIFICATE OF THE SUPERVISOR

This is to certify that the thesis entitled “*Development of composite conducting polymer with 2D layered and metallic nanosystems for biofunctionalization and electrochemical biosensing applications*” submitted to the School of Sciences, Tezpur University in partial fulfilment for the award of the degree of Doctor of Philosophy in Physics, is a record research work carried out by **Ms. Susmita Baruah** under my supervision and guidance.

All help received by her from various sources has been duly acknowledged. No part of this thesis has been submitted elsewhere for the award of any other degree.

Place: *Tezpur, Assam*

Date: *31/07/2025.*

(Prof. Dambarudhar Mohanta)

Designation: Professor
School: School of Sciences
Department: Physics
Tezpur University
Tezpur-784028, Assam, India

Acknowledgements

First and foremost, I would like to express my heartfelt gratitude to my late supervisor, Prof. Ashok Kumar, for his unwavering belief in my abilities and invaluable guidance during my time as a Junior Research Fellow (JRF). Though he is no longer with us, his inspiration and support continue to guide me, and his lasting impact will always be cherished.

It is with immense gratitude and deep appreciation that I acknowledge the invaluable guidance of my esteemed supervisor, Prof. Dambarudhar Mohanta, whose dynamic leadership, meticulous attention to detail, and unwavering support have been influential throughout my Ph.D. journey. I will forever be grateful for his expertise, patience, and dedication, which have been crucial in transforming challenges into opportunities for learning and growth.

I would like to extend my heartfelt appreciation to Dr. C. A. Betty, Scientist at BARC. I consider myself extremely fortunate to have had the privilege of her mentorship. Dr. Betty's guidance, patience, and expertise significantly contributed to shaping my research and broadened my perspective in ways that would not have been possible otherwise. Her constructive feedback, combined with her vast knowledge, has been an essential part of my academic and professional growth, and I will always be thankful for the many lessons she imparted to me. Her influence will remain a lasting source of inspiration as I continue my academic and professional endeavours.

I would like to express my sincere gratitude to my DRC members, Prof. Pabitra Nath, Dr. Shyamal Kumar Das and Prof. Pralay Kumar Karmakar from the Department of Physics, for their continuous motivation, valuable suggestions, and insightful comments throughout my Ph.D. work. I am highly grateful to Prof. Mrinal Kumar Das for his supports and encouragement during my tough time.

I am deeply grateful to the Board of Research in Nuclear Sciences (BRNS), Mumbai, for their generous financial support through the project (project number-58/14/09/2019-BRNS).

My earnest gratitude to all the members of Materials Research Laboratory (MRL), Department of Physics with whom I got the opportunity to work and share our knowledge; Rituraj da, Bhagyalakhi ba, Devalina ba, Kashmiri ba and Atowar da. I would also like to extend my thanks to both the past and present members of the Nanoscience and Soft-Matter Laboratory (NSL), Department of Physics, for their continuous support, assistance, and motivation throughout my research; Kakoli ba, Ankush da, Aftab da, Jimly, Stuti, Mahesh, Anuj, Bikash, Dipankar, Saddique, Satyajit and Risa. I consider myself fortunate to have had such wonderful labmates, whose encouragement and company have been invaluable. I would also to express my thanks to my friends Dhanjit, Shahnaz, Sanghu ba, Priyanka ba, Asha ba and Kasturi; for making my time at Tezpur University enjoyable and memorable. I would like to express my heartfelt gratitude to all my friends specially Nihu, Pori, Dhana, Tris, and Ilu who have been an integral part of my academic and personal life during this journey.

I would like to acknowledge the technical assistants of Department of Physics and SAIC, Biju da, Tridib da, Mrinmoy da, and Sankur da for helping me to carry out characterizations. I would offer my sincere gratitude to Narayan da and Patir da for all the help provided to complete the office works. I would like to thank Sophisticated analytical instrument facility (SAIF), Gauhati University for providing the instrumentation facilities. I would like to express my sincere appreciation to SAIC, Tezpur University, as well as UGC-SAP DRS II and DST-FIST programs of the Department of Physics, Tezpur University, for their valuable support in terms of providing resources, facilitating the execution of ideas.

I would like to express my deepest gratitude to my parents, whose unconditional love, support, and sacrifices have been the foundation of my academic journey. To my mother, your boundless love and nurturing care have been a constant source of strength and encouragement throughout my life. Your selflessness and devotion have provided me with the emotional and moral support I needed during the most challenging moments of my Ph.D. Your belief in me and almighty has always been a guiding light, and your encouragement has helped me overcome every obstacle with confidence and determination. To my dangor ma, your companionship, care, love, and understanding have provided me with the strength to face the challenges that came with this long academic journey. Your belief in me has been a pillar of support that I will forever cherish.

I would like to express my heartfelt tribute to my late father. His untimely passing left a void, but also a profound sense of motivation to honor his memory by striving for the very best in everything I do. I dedicate this thesis to him, in gratitude for the foundation he gave me in the early years of my life.

I am also profoundly grateful to my brother for his constant encouragement and support. From childhood, he has been my protector, guide, and confidant. His selflessness and dedication in looking after me have shaped me into the person I am today. As a role model and a source of strength, he has always been there to encourage me, especially during the most challenging times of my journey. I am truly fortunate to have had such a loving and supportive brother, whose presence has always made my path easier and my burdens lighter. I am Thankful to Hira ba as well for all the love, support and care. Together, the love, sacrifices, and encouragement of my family have been the cornerstone of my success. Without their unwavering presence, understanding, and guidance, this achievement would not have been possible. I will forever be grateful for everything they have done for me.

I would like to extend my deepest gratitude to Swarnav whose constant support has been the cornerstone of my PhD journey. You stood by me through every challenge, celebrating my victories and lifting me up in times of struggle, reminding me to keep pushing forward. Thank you for being my greatest source of comfort throughout this journey. Thank you for being my confidant and my source of endless inspiration. This milestone is as much yours as it is mine, and I am forever grateful for your presence in my life.

Lastly, I would like to express my deepest gratitude to the Almighty for showering countless blessings upon me throughout this journey. At every step, the guidance and grace have been my source of strength, providing me with the resilience to overcome challenges and the clarity to move forward.

Date: 31/07/2025

Place: Tezpur

Susmita Baruah
(Susmita Baruah)

List of Tables

Chapter	Table no.	Caption	Page no.
Chapter 2	2.1	Physical properties of monomer	52
	2.2	Physical properties of functionalizing agents	53
	2.3	Physical properties of oxidizing agents and solvents	53
	2.4	Physical properties of the biofunctionalization element, blocking agent and enzyme	54
	2.5	Physical properties of the analytes	54
Chapter 3	3.1	Atomic percentage of the elemental constituents for; (a) PEDOT-PSS, (b) GO/PEDOT-PSS and (c) AuNP/GO/PEDOT-PSS.	101
	3.2	Atomic percentage of the elemental constituents for (a) WS ₂ /PEDOT-PSS and (b) AuNP/WS ₂ /PEDOT-PSS.	103
	3.3	EIS spectra fitted parameters of PEDOT-PSS/ITO, GO/PEDOT-PSS/ITO and AuNP/GO/PEDOT-PSS/ITO electrodes	108
	3.4	EIS spectra fitted parameters of WS ₂ /PEDOT-PSS/ITO and AuNP/WS ₂ /PEDOT-PSS/ITO electrodes.	109
	3.5	Calculated values of electroactive area	110
Chapter 4	4.1	Fitted parameters of EIS spectra for the ternary (AuNP/GO/PEDOT-PSS) electrode after every process step of sensor fabrication	125
	4.2	Results of AuNPs embedded GO/PEDOT-PSS glucose sensor from repeatable experiment	128
	4.3	Fitted parameters of impedance spectra for AuNP/GO/PEDOT-PSS/ITO electrode after every process step and after addition of AF-B ₁	133
	4.4	Results from repeated experiments for mouse IgG/GLU/AuNP/GO/PEDOT-PSS capacitive mouse derived anti-AF-B ₁ sensor (both at 77 Hz and 1 kHz)	136

	4.5	Results from repeated experiments for AF-B ₁ capacitive sensor (both at 77 Hz and 1 kHz)	140
	4.6	Results from repeated experiments for anti-AF-B ₁ /GLU/AuNP/GO/PEDOT-PSS based AF-B ₁ sensor using DPV	143
Chapter 5	5.1	Fitted parameters from EIS spectra for AuNPs deposited WS ₂ /PEDOT-PSS/ITO electrode after every process step of sensor fabrication	160
	5.2	Results of AuNP/WS ₂ /PEDOT-PSS glucose sensor from repeatable experiment	165
	5.3	Fitted parameters of impedance spectra for AuNP/WS ₂ /PEDOT-PSS/ITO electrode after every process step and after addition of AF-B ₁	171
	5.4	Results from repeated experiments for AF-B ₁ capacitive sensor (both at 77 Hz and 1 kHz)	174
	5.5	Results of repeated experiments for Au/WS ₂ /PEDOT-PSS based AF-B ₁ sensor using amperometry technique, (DPV)	177
Chapter 6	6.1	Fitted parameters from EIS spectra for PANI-PVA/ITO electrode after every process step of sensor fabrication	189
	6.2	Results of PANI-PVA glucose sensor from repeatable experiment	193
	6.3	Fitted parameters of impedance spectra for PANI-PVA/ITO electrode after every process step and after addition of AF-B ₁	199
	6.4	Results from repeated experiments for AF-B ₁ capacitive sensor (both at 77 Hz and 1 kHz)	203
	6.5	Repeated experiments result for PANI-PVA based AF-B ₁ sensor using amperometry technique, DPV	206
Chapter 7	7.1	Detection of glucose was carried out via chronoamperometry	218
	7.2	Detection of Aflatoxin B ₁ carried out via Transient capacitance and DPV	219

List of Figures

Chapter	Figure no.	Caption	Page no.
Chapter 1	1.1	Diagrammatic representation of band gap in insulator, semiconductor and metal	3
	1.2	Schematic representation of band formation in conducting polymers	4
	1.3	Chemical structure of PEDOT-PSS	6
	1.4	Structure of Polyaniline	7
	1.5	Structure of different oxidation states of polyaniline	8
	1.6	Schematic structure of GO	10
	1.7	Schematic layered structure of WS ₂	11
	1.8	Schematic representation of an electrochemical biosensor setup	14
	1.9	Schematic representation of interaction between enzyme and substrate	19
	1.10	Schematic representation of an antibody	20
	1.11	Schematic representation of interaction between antibody with the target analyte	21
	1.12	The diagrammatic representation of (a) adsorption, (b) covalent bonding, (c) cross linking, (d) entrapment, and (e) encapsulation of immobilization of enzyme	22
Chapter 2	2.1	Schematic illustration for the preparation of PEDOT-PSS/ITO electrode	55
	2.2	Deposition curve of PEDOT-PSS film on ITO at a scan rate of 20 mV/s	55
	2.3	Schematic illustration for synthesis of graphene oxide	56
	2.4	Schematic illustration for preparation of GO/PEDOT-PSS/ITO electrode	57
	2.5	Deposition curve of GO/PEDOT-PSS film on ITO at a scan rate of 20 mV/s	57

2.6	Schematic illustration of preparation of Au/GO/PEDOT-PSS/ITO electrode	58
2.7	Deposition curves of AuNP/GO/PEDOT-PSS films on ITO at a scan rate of 20 mV/s	59
2.8	Schematic illustration of synthesis of WS ₂ nanosheets	60
2.9	Schematic illustration of preparation of WS ₂ /PEDOT-PSS/ITO electrode	60
2.10	Deposition curves of WS ₂ /PEDOT-PSS/ITO electrode within the potential window from – 0.2 V to 1.2 V	61
2.11	Schematic illustration of preparation of Au/WS ₂ /PEDOT-PSS/ITO electrode	61
2.12	Deposition curves of Au/WS ₂ /PEDOT-PSS/ITO electrode within the potential window from – 0.2 V to 1.2 V	62
2.13	Schematic illustration of chemical synthesis of polyaniline (PANI)	64
2.14	Schematic illustration of preparation of PANI nanofibers	65
2.15	Schematic illustration of preparation of PANI-PVA composite solution	65
2.16	Schematic illustration of preparation of PANI-PVA nanofibers	66
2.17	Electrospinning unit during experiment	66
2.18	Schematic illustration of the AuNP anchored 2D (GO or WS ₂) modified PEDOT-PSS/ITO based glucose sensor	67
2.19	Schematic illustration of the PANI-PVA nanofiber based glucose sensor	68
2.20	Schematic of electrochemical based immunosensor (AuNP/2D/PEDOT-PSS) fabrication and testing	69

	2.21	Schematic of electrochemical based immunosensor (PANI-PVA) fabrication and testing	69
	2.22	Triangular potential waveform used in cyclic voltammetry showing the variation of potential from an initial value (E1) to a final value (E2)	75
	2.23	The cyclic voltammogram of a Nernstian electrochemical reaction	76
	2.24	(i) The input sinusoidal potential and (ii) corresponding current signal in output of the electrochemical cell	78
	2.25	(a) Nyquist plot and (b) corresponding simplified Randles equivalent circuit	80
	2.26	(a) Nyquist plot for a mixed circuit and (b) corresponding Randles equivalent circuit	81
Chapter 3	3.1	XRD spectra of PEDOT-PSS, GO/PEDOT-PSS and AuNP/GO/PEDOT-PSS films over ITO electrode	94
	3.2	XRD spectra of PEDOT-PSS, WS ₂ /PEDOT-PSS and AuNP/WS ₂ /PEDOT-PSS films over ITO electrode	94
	3.3	XRD plot of pure PANI, PANI/PVA composite fibers	95
	3.4	FTIR spectra of PEDOT-PSS, GO/PEDOT-PSS and AuNP/GO/PEDOT-PSS films over ITO electrode	96
	3.5	FTIR spectra of synthesized PEDOT-PSS, WS ₂ /PEDOT-PSS, and AuNP/WS ₂ /PEDOT-PSS films	98
	3.6	FTIR analysis of pure PANI, PANI/PVA composite fibers	99
	3.7	FESEM images of (a) PEDOT-PSS, (c) GO/PEDOT-PSS and (e) AuNP/GO/PEDOT-PSS films on ITO. (g, h) magnified view of Fig. (e). EDX analysis of (b) PEDOT-PSS, (d) GO/PEDOT-PSS and (f) AuNP/GO/PEDOT-PSS films on ITO	100
	3.8	FESEM images of (a) WS ₂ /PEDOT-PSS and (c) AuNP/WS ₂ /PEDOT-PSS films on ITO. (e, f)	102

	magnified view of Fig. (c). EDX spectra of (b) WS ₂ /PEDOT-PSS and (d) AuNP/WS ₂ /PEDOT-PSS films on ITO	
3.9	FESEM images of (a) PANI particles and PANI-PVA nanofibers (b) 1:1 ratio (c) 1:2 ratio (d) 1:3 ratio. (e,f) magnified view of (d)	104
3.10	Cyclic voltammetry plot of (a) PEDOT-PSS, GO/PEDOT-PSS, WS ₂ /PEDOT-PSS, AuNP/GO/PEDOT-PSS and AuNP/WS ₂ /PEDOT-PSS films on ITO; (b) PEDOT-PSS/ITO, GO/PEDOT-PSS/ITO, Au/GO/PEDOT-PSS/ITO (c) PEDOT-PSS, WS ₂ /PEDOT-PSS and Au/WS ₂ /PEDOT-PSS/ITO	106
3.11	Nyquist plot of (a) PEDOT-PSS, GO/PEDOT-PSS, AuNP/GO/PEDOT-PSS (b) WS ₂ /PEDOT-PSS, AuNP/WS ₂ /PEDOT-PSS films on ITO.	107
3.12	CV plots recorded with scan rate variation for ITO electrode modified with (a) PEDOT-PSS, (b) GO/PEDOT-PSS, (c) WS ₂ /PEDOT-PSS, (d) AuNP/GO/PEDOT-PSS, (e) AuNP/WS ₂ /PEDOT-PSS, (f) PANI-PVA. (inset: i_p vs $v^{1/2}$ linear plot for respective electrodes, S is slope)	111
Chapter 4		
4.1	Cyclic voltammetry plots of (a) PEDOT-PSS/ITO sensor (with 10 μ M glucose), (b) GO/PEDOT-PSS/ITO sensor after each process step, with final step of 13.62 μ M glucose interaction, (c) AuNP/GO/PEDOT-PSS/ITO sensor after each process step and with addition of 347.72 μ M glucose, (d) <i>Glox</i> /GLU/AuNP/GO/PEDOT-PSS bioelectrode at different glucose concentrations	122
4.2	Impedance response of (a) PEDOT-PSS/ITO sensor (glucose concentration 10 μ M), (b) GO/PEDOT-PSS/ITO sensor (with addition of glucose concentration 13.62 μ M), (c) AuNP/GO/PEDOT-	123

	PSS/ITO sensor (with addition of glucose concentration 347.72 μM) after each process steps	
4.3	Chronoamperometry plot of (a) PEDOT-PSS/ITO (b) GO/PEDOT-PSS/ITO electrodes at different glucose concentrations at a potential of 0.8 V	126
4.4	(a) Chronoamperometry plot of AuNP/GO/PEDOT-PSS/ITO electrode at different concentrations of glucose. (b) Selectivity test in presence of different interference like ascorbic acid (AA), uric acid (UA), sucrose (1 mM of each). (c) Linear regression plot of current vs. concentration for AuNP/GO/PEDOT-PSS/ITO biosensor with error bar. (d) Linear regression plot of Conc. (both μM and mg/dL) vs ΔI for AuNP/GO/PEDOT-PSS/ITO	127
4.5	Chronoamperometry plots of AuNP/GO/PEDOT-PSS/ITO electrode at different volume of salivary sample (a) sample 1, (b) sample 2. Linear regression plot of current vs. volume for AuNP/GO/PEDOT-PSS/ITO electrode of (c) sample 1, (d) sample 2 with error bars. Linear regression plot of Conc. (both μM and mg/dL) vs ΔI (e) sample 1 and (f) sample 2	129
4.6	Cyclic voltammetry plot of anti-AF-B ₁ /AuNP/GO/PEDOT-PSS/ITO sensor within a potential window of -0.2 V to 1.2 V and at a scan rate of 20 mV/s in (a) PBS (electrolyte1) after each process step, with final step of injecting antigen concentration. (b) 100 mM PBS and 5 mM $[\text{Fe}(\text{CN})_6]^{3-/4-}$ (electrolyte 2) after 594.59 ng mL^{-1} AF-B ₁ addition	131
4.7	Impedance plot of AuNP/GO/PEDOT-PSS films on ITO electrode after each process step and antigen concentration as a final step in PBS electrolyte. The inset shows the equivalent circuit and high frequency impedance enlarged	132

4.8	Transient capacitance plot of mouse IgG/GLU/AuNP/GO/PEDOT-PSS based immunosensor at (a) 77 Hz, (b) 1 kHz towards detection of anti-AF-B ₁ raised in mouse. (c) Langmuir plot of mouse IgG/GLU/AuNP/GO/PEDOT-PSS sensors at 77 Hz and 1 kHz. (d) Linear regression plot of mouse IgG/GLU/AuNP/GO/PEDOT-PSS capacitive sensor at 77 Hz and 1 kHz	135
4.9	Linear regression plot of mouse IgG/GLU/AuNP/GO/PEDOT-PSS capacitive sensors at 77 Hz and 1 kHz with error bar towards detection of mouse derived anti-AF-B ₁	137
4.10	Transient capacitance plot of anti-AF-B ₁ /GLU/AuNP/GO/PEDOT-PSS based AF-B ₁ sensors at (a) 77 Hz at a DC bias voltage of 0.8 V (b) Selectivity test. (c) Adsorption plot and (d) Linear regression plot of anti-AF-B ₁ /GLU/AuNP/GO/PEDOT-PSS capacitive sensor at 77 Hz and 1 kHz	138
4.11	Transient capacitance plot of anti-AF-B ₁ /GLU/AuNP/GO/PEDOT-PSS based AF-B ₁ sensors at 1 kHz	139
4.12	Linear regression plot of anti-AF-B ₁ /GLU/AuNP/GO/PEDOT-PSS capacitive sensors at 77 Hz and 1 kHz with error bar	140
4.13	DPV plot of BSA/anti-AF-B ₁ /GLU/AuNPs/GO/PEDOT-PSS/ITO with concentration of AF-B ₁ (18.18 ng mL ⁻¹ -594.59 ng mL ⁻¹) within potential window -0.2 V to 0.8 V. (b) Selectivity test. (c) Adsorption isotherm fitted for change in peak current of DPV with change in antigen concentration. (d)	142

	Linear regression fitting for $1/\Delta\text{Conc.}$ vs. $1/\Delta\text{Current}$ from DPV response	
4.14	$1/\Delta I$ vs. $1/\Delta\text{Conc}$ for the BSA/anit-AF-B ₁ /GLU/AuNP/GO/PEDOT-PSS sensor along with linear regression plot with error bar w.r.to standard deviation	143
4.15	Scheme for testing white button Mushroom, Okra of different days of air exposure at 20-24 °C after purchase	144
4.16	SEM images of (a) as purchased mushroom, (b) 3 days stored mushroom, (c) as purchased okra (d) 7 days stored okra	145
4.17	DPV pattern of AuNP/GO/PEDOT-PSS/ITO composite electrode towards mushrooms (kept exposed to air 20 °C for 0-3 days and washed with PBS) with different spiking volumes (50-200) μL , (a) 50 μL (b) 100 μL , (c) 200 μL . Inset shows the voltammogram after addition of 200 μL PBS	146
4.18	DPV pattern of AuNP/GO/PEDOT-PSS/ITO composite electrode towards mushrooms (kept exposed to air 20 °C for 0-3 days and washed with PBS) with spiking volumes 300 μL . Inset shows the voltammogram after addition of 200 μL PBS. (b) Change in peak current vs number of days of air exposure for mushrooms	146
4.19	DPV pattern of AuNP/GO/PEDOT-PSS/ITO composite electrode towards okra (kept exposed to air 24 °C for 0-7 days and washed with PBS) with different spiking volumes (50-200) μL , (a) 50 μL (b) 100 μL , (c) 200 μL . Inset of each figure shows the voltammogram after addition of 200 μL PBS	147
4.20	(a) DPV pattern of AuNP/GO/PEDOT-PSS/ITO composite electrode towards okra (kept exposed to air	147

		24 °C for 0-7 days after purchase and washed with PBS) with spiking volumes 300 μ L. Inset shows the voltammogram after addition of 200 μ L PBS. (b) Change in peak current vs number of days of air exposure for okra	
Chapter 5	5.1	Cyclic voltammetry plots of (a) <i>G/Ox</i> /GLU/WS ₂ /PEDOT-PSS/ITO sensor after each process step, with final step of 32 μ M glucose interaction, (b) <i>G/Ox</i> /GLU/AuNP/WS ₂ /PEDOT-PSS/ITO sensor after each process step and with addition of 440.67 μ M glucose	158
	5.2	Impedance response of (a) WS ₂ /PEDOT-PSS/ITO sensor (with addition of glucose concentration 32 μ M), (b) AuNP/WS ₂ /PEDOT-PSS/ITO sensor (with addition of glucose concentration 440.67 μ M) after each process steps	161
	5.3	Cyclic voltammetry plot of <i>G/Ox</i> /GLU/AuNP/WS ₂ /PEDOT-PSS bioelectrode at different glucose concentrations	162
	5.4	Chronoamperometry plot of WS ₂ /PEDOT-PSS/ITO electrode at different glucose concentrations in PBS electrolyte	163
	5.5	Chronoamperometry plots of AuNP/WS ₂ /PEDOT-PSS/ITO electrode at different concentrations of glucose. (b) Linear regression plot of current vs. concentration for AuNP/WS ₂ /PEDOT-PSS/ITO biosensor with error bar. (c) Linear regression plot of Conc. (both μ M and mg/dL) vs ΔI for AuNP/WS ₂ /PEDOT-PSS/ITO	164
	5.6	Selectivity test in presence of different interference like ascorbic acid (AA), uric acid (UA), sucrose (1 mM of each)	165

5.7	Chronoamperometry plots of AuNP/WS ₂ /PEDOT-PSS/ITO electrode at different volume of salivary sample (a) sample 1, (b) sample 2. Linear regression plot of current vs. volume for AuNP/WS ₂ /PEDOT-PSS/ITO electrode of (c) sample 1, (d) sample 2 with error bars. Linear regression plot of Conc. (both μM and mg/dL) vs ΔI (e) sample 1 and (f) sample 2	166
5.8	Cyclic voltammetry plot of (a) anti-AF-B ₁ /WS ₂ /PEDOT-PSS/ITO within a potential window of -0.2 V to 1.2 V and at a scan rate of 20 mV/s. CV plots of anti-AF-B ₁ /AuNP/WS ₂ /PEDOT-PSS/ITO sensor in (b) PBS (electrolyte1) after each process step, with final step of injecting antigen concentration. (c) 100 mM PBS and 5 mM $[\text{Fe}(\text{CN})_6]^{3-/4-}$ (electrolyte 2) after AF-B ₁ addition	168
5.9	Impedance plot of (a) WS ₂ /PEDOT-PSS and (b) AuNP/WS ₂ /PEDOT-PSS films on ITO electrode after each process. The inset shows the equivalent circuit	170
5.10	Transient capacitance plot of anti-AF-B ₁ /GLU /Au/ WS ₂ /PEDOT-PSS based AF-B ₁ sensors at (a) 77 Hz (b) 1 kHz; at a DC bias voltage of 0.8 V. (c) Adsorption plot and (d) Linear regression plot of Au/WS ₂ /PEDOT-PSS capacitive sensor at 77 Hz and 1 kHz	172
5.11	Linear regression plot of AuNP/WS ₂ /PEDOT-PSS capacitive sensors at 77 Hz and 1 kHz; with error bar	174
5.12	DPV plot of BSA/anti-AF-B ₁ /GLU/Au-NPs/WS ₂ /PEDOT-PSS/ITO with different concentration of AF-B ₁ (7.69 ng mL ⁻¹ -654.25 ng mL ⁻¹) (b) Adsorption isotherm plot. (c) Linear regression plot for $1/\Delta\text{Conc.}$ vs. $1/\Delta\text{Current}$	175
5.13	$1/\Delta I$ vs. $1/\Delta\text{Conc.}$ for the BSA/anit-AF-B ₁ /GLU/AuNP/WS ₂ /PEDOT-PSS sensor along with	176

	linear regression plot with error bar w.r.to standard deviation.	
5.14	Selectivity test of BSA/anit-AF-B ₁ / GLU/AuNP/ WS ₂ / PEDOT-PSS sensor	177
5.15	Scheme for testing peanuts of different days of air exposure at a temperature of 27 °C after obtaining from local source	178
5.16	DPV pattern of AuNP/WS ₂ /PEDOT-PSS/ITO composite electrode towards peanuts (kept exposed to air 27 °C for 0-25 days and washed with PBS) with different spiking volumes (50-200) µL, (a) 50 µL (b) 100 µL, (c) 200 µL (d) 300 µL. Inset shows the voltammogram after addition of 200 µL PBS	179
5.17	Change in peak current vs number of days of air exposure for peanuts	180
Chapter 6		
6.1	Cyclic voltammetry plots of <i>Glo</i> x/GLU/PANI-PVA/ITO sensor after each process step, with final step of 460.87 µM glucose interaction	188
6.2	Impedance response of PANI-PVA/ITO sensor (with addition of glucose concentration 460.87 µM) after each process steps	189
6.3	Cyclic voltammetry plot of <i>Glo</i> x/GLU/PANI-PVA bioelectrode at different glucose concentrations	190
6.4	(a) Chronoamperometry plots of PANI-PVA/ITO electrode at different concentrations of glucose. (b) Linear regression plot of current vs. concentration for PANI-PVA /ITO biosensor with error bar. (c) Linear regression plot of Conc. (both µM and mg/dL) vs ΔI for PANI-PVA /ITO	192
6.5	Selectivity test of PANI-PVA sensor in presence of different interference like ascorbic acid (AA), uric acid (UA), sucrose (1 mM of each)	193

- 6.6 Chronoamperometry plots of PANI-PVA/ITO electrode at different volume of salivary sample (a) sample 1, (b) sample 2. Linear regression plot of current vs. volume for PANI-PVA/ITO electrode of (c) sample 1, (d) sample 2 with error bars. Linear regression plot of Conc. (both μM and mg/dL) vs ΔI (e) sample 1 and (f) sample 2 195
- 6.7 Cyclic voltammetry plot of anti-AF-B₁/PANI-PVA/ITO at a scan rate of 20 mV/s (a) within a potential window of -0.4 V to 1.0 V in PBS (electrolyte1) after each process step, with final step of injecting antigen concentration (b) within a potential window of -0.4 V to 0.8 V in 100 mM PBS and 5 mM $[\text{Fe}(\text{CN})_6]^{3-/4-}$ (electrolyte 2) after AF-B₁ addition 197
- 6.8 Impedance plot of PANI-PVA films on ITO electrode after each process. The inset shows the equivalent circuit 199
- 6.9 Transient capacitance plot of anti-AF-B₁/GLU/PANI-PVA based AF-B₁ sensors at (a) 77 Hz (b) 1 kHz. (c) Adsorption plot and (d) Linear regression plot of PANI-PVA capacitive sensor at 77 Hz and 1 kHz 201
- 6.10 Linear regression plot of PANI-PVA capacitive sensors at 77 Hz and 1 kHz; with error bar 203
- 6.11 DPV plot of BSA/anti-AF-B₁/GLU/PANI-PVA/ITO with different concentration of AF-B₁ (3.92 ng mL^{-1} - 690 ng mL^{-1}) (b) Adsorption isotherm plot. (c) Linear regression plot for $1/\Delta\text{Conc.}$ vs. $1/\Delta\text{Current}$ 204
- 6.12 $1/\Delta I$ vs. $1/\Delta\text{Conc.}$ for the BSA/anti-AF-B₁/GLU/PANI-PVA sensor along with linear regression plot with error bar w.r.to standard deviation 206

6.13	Selectivity test of BSA/anit-AF-B ₁ /GLU/PANI-PVA sensor	207
6.14	DPV pattern of PANI-PVA/ITO composite electrode towards mushrooms (kept exposed to air 20 °C for 0-3 days and washed with PBS) with different spiking volumes (50-200) μ L, (a) 50 μ L (b) 100 μ L, (c) 200 μ L (d) 300 μ L. Inset shows the voltammogram after addition of 200 μ L PBS	208
6.15	Change in peak current vs number of days of air exposure for mushrooms	209
6.16	DPV pattern of PANI-PVA/ITO composite electrode towards okra (kept exposed to air 24 °C for 0-7 days and washed with PBS) with different spiking volumes (50-200) μ L, (a) 50 μ L (b) 100 μ L, (c) 200 μ L (d) 300 μ L. Inset of each figure shows the voltammogram after addition of 200 μ L PBS	209
6.17	Change in peak current vs number of days of air exposure for okra	210

List of abbreviations

HOMO	: Highest occupied molecular orbital
LUMO	: Lowest unoccupied molecular orbital
PEDOT-PSS	: Poly(3,4-ethylenedioxythiophene)-polystyrene sulfonate
PANI	: Polyaniline
GO	: Graphene oxide
TMDCs	: Transition metal dichalcogenides
WS ₂	: Tungsten disulfide
AuNP	: Gold nanoparticle
PVA	: Polyvinyl alcohol
NMP	: N-Methyl-2-Pyrrolidone
ITO	: Indium tin oxide
IgG	: Immunoglobulins G
BSA	: Bovine serum albumin
GLU	: Glutaraldehyde
<i>Glox</i>	: Glucose Oxidase
AF-B ₁	: Aflatoxin B ₁
XRD	: X-ray diffraction
FTIR	: Fourier Transform Infrared Spectroscopy
SEM	: Scanning electron microscopy
EDX	: Energy dispersive x-ray
CV	: Cyclic voltammetry
EIS	: Electrochemical Impedance spectroscopy
CPE	Constant Phase Element
PBS	: Phosphate buffer saline
DPV	: Differential pulse voltammetry
kD	: Kilo-Dalton
POC	: Point-of-care
FAD	: Flavin adenine dinucleotide
LOD	: Limit of detection

List of symbols

R_l	: Solution resistance
R_{ct}	: Charge transfer resistance
C_{dl}	: Double layer capacitance
ω	: Warburg impedance
Q	: Constant phase element
Z'	: Real part of impedance
Z''	: Imaginary part of impedance
σ	: Standard deviation
m	: Slope
v	: Scan rate
D	: Diffusion coefficient
A	: Electroactive area
C	: Concentration
i_p	: Anodic peak current
i_c	: Cathodic peak current
C_p	: Capacitance